



CERTIFICATE OF COMPLIANCE
LLOYD'S REGISTER NEDERLAND B.V.

hereby declares that the certification assessment has demonstrated that
**foodproof® *Salmonella* spp. method 5'Nuclease or hybridization probes (liquid
and lyophilized), automated protocol**

Manufactured and supplied by:
BIOTECON Diagnostics GmbH
Hermannswerder 17
14473 Potsdam
Germany

The foodproof *Salmonella* Detection (Lyo)Kits R 310 27, R 302 27 and R 602 27 in combination with the foodproof Magnetic Preparation Kit I S 400 11 L have been validated and revealed to be at least equivalent to the reference method as demonstrated by the validation study report. The summary of the validation report is available on the MicroVal website: www.microval.org

Reference method: ISO 6579-1:2017 Microbiology of the food chain - Horizontal method for the detection, enumeration and serotyping of *Salmonella* - Part 1: Detection of *Salmonella* spp.

Scope

- Liquid kit: chocolate and bakery products, meat and meat products, milk and dairy products, egg products, raw beef meat 375 g and feed samples;
- Lyophilized kit: broad range of foods (chocolate and bakery products, meat and meat products, milk and dairy products, egg products, fish and seafood products) and feed samples

The validation and certification has been performed in accordance with EN ISO 16140-2:2016 and the MicroVal Rules and Certification Scheme version 8.


Certificate no.: 2011LR40

First approval date: 26 June 2013

Renewal date: 25 June 2021

Expiry date: 25 June 2025

ISSUED BY:



Lloyd's Register Nederland B.V.
Rotterdam, The Netherlands