

CERTIFICATE OF COMPLIANCE

LRQA

hereby declares that the certification assessment has demonstrated that
**foodproof® *Enterobacteriaceae* plus *Salmonella* Detection LyoKit –
5´Nuclease –
used with thermocycler LightCycler® 480**

Manufactured and supplied by:
BIOTECON Diagnostics GmbH
Hermannswerder 17
14473 Potsdam
Germany

has been validated and revealed to be at least equivalent to the reference method as demonstrated by the validation study report. The summary of the validation report is available on the MicroVal website: www.microval.org

Reference methods:

- ISO 21528-1:2017 Microbiology of the food chain — Horizontal method for the detection and enumeration of *Enterobacteriaceae* — Part 1: Detection of *Enterobacteriaceae*
- ISO 6579-1:2017 Microbiology of the food chain - Horizontal method for the detection, enumeration and serotyping of *Salmonella* - Part 1: detection of *Salmonella* spp.
- ISO 6579-1:2017/A1:2020 Microbiology of the food chain – Horizontal method for the detection, enumeration and serotyping of *Salmonella* spp. - Part 1: detection of *Salmonella* spp. Amendment 1: Broader range of incubation temperatures, amendment to the status of Annex D, and correction of the composition of MSRV and SC.

Scope: infant cereals, infant formula with or without probiotics and ingredients, production environmental samples

The validation and certification has been performed in accordance with EN ISO 16140-2:2016 and the MicroVal Rules and Certification Scheme version 8.

Certificate no.: 2020LR90

First approval date: 10 March 2022

Expiry date: 9 March 2026



ISSUED BY: _____
LRQA Nederland B.V.
Rotterdam, The Netherlands