



## CERTIFICATE OF COMPLIANCE LRQA

hereby declares that the certification assessment has demonstrated that

## foodproof® Salmonella spp. method (liquid and lyophilized) and vetproof® Salmonella spp. method (lyophilized), manual protocol

Manufactured and supplied by:
Hygiena Diagnostics GmbH
Hermannswerder 17
14473 Potsdam
Germany

The foodproof® Salmonella Detection (Lyo)Kits R 310 27, R 302 27 and R 602 27 and the vetproof® Salmonella qPCR LyoKit (also know as Salmonella Genus DNA Detection Test Kit V 900 27) in combination with the foodproof® StarPrep One Kit S 400 07 have been validated and revealed to be at least equivalent to the reference method as demonstrated by the validation study report, which has been prepared by a MicroVal Expert Laboratory and approved by the MicroVal Technical Committee. The above-mentioned production facility complies to the QMS requirements as mentioned in the MicroVal Rules. The summary of the validation report is available on the MicroVal website: <a href="www.microval.org">www.microval.org</a>.

Reference method: ISO 6579-1:2017- Microbiology of the food chain - Horizontal method for the detection, enumeration and serotyping of Salmonella - Part 1: Detection of *Salmonella* spp.

## Scope:

- Liquid kit: chocolate and bakery products, meat and meat products, milk and dairy products, egg products, raw beef meat 375 g, feed samples, primary production samples;
- Lyophilized kit: broad range of foods (chocolate and bakery products, meat and meat products, milk and dairy products, egg products, fish and seafood products), feed samples and primary production samples.

The validation and certification has been performed in accordance with EN ISO 16140-2:2016 and the MicroVal Rules.

Certificate no.: 2011LR42 First approval date: 26 June 2013
Renewal date: 19 June 2025

Expiry date: 25 June 2029

ISSUED BY: LRQA Nederland B.V.

Rotterdam, The Netherlands

Certificate no.: 2011LR42 20-06-2025 Page 1 of 1