



## CERTIFICATE OF CONFORMITY

EnSURE™ Touch, portable palm-sized luminometers, are manufactured under a ISO 9000 Quality System and comply with the following standards.

General	<ul style="list-style-type: none"><li>• Low Voltage Directive 2014/35/EU</li><li>• EMC Directive 2014/30/EU</li><li>• RoHS Directive 2011/65/EU</li><li>• PAH (Polycyclic Aromatic Hydrocarbons)</li></ul>
Safety	<ul style="list-style-type: none"><li>• BS EN 61010-1:2010, IEC 61010-1:2010</li><li>• UL 61010-1:2012 R4.16</li><li>• CAN/CSA C22.2 NO. 61010.1-12 + GI1 + GI2 (R2017)</li></ul>
Emissions	<ul style="list-style-type: none"><li>• ETSI EN 301 489-1 V2.1.1, ETSI EN 301 489-17 V3.1.1</li><li>• EN 55011:2009 + A1:2010, FCC Part 15 Subpart B:2018</li><li>• AS/NZ CISPR11:2015 +A1:2016 +A2:2019, ICES-003 Issue 6</li><li>• EN 61000-3-2:2014, EN 61000-3-3:2013</li></ul>
Immunity	<ul style="list-style-type: none"><li>• EN 61326-1:2013</li><li>• EN 61000-4-2:2009, EN 61000-4-3:2006 + A1:2018 + A2:2010</li><li>• EN 61000-4-4:2012, EN 61000-4-5:2006, EN 61000-4-6:2009</li><li>• EN 61000-4-8:2010, EN 61000-4-11:2004</li></ul>

The EnSURE Touch Luminometer with serial number XXXXXX has been tested and performs in accordance with the product quality control specifications.

### Positive Control

B Light calibration range 255 - 345 RLU

CV's @ calibration point +/- 5%

### Negative Control

Background (Blank) reading range 0 - 5 RLU

---

QC signature:

Date: